

## Instrument Response Information

Broadband Sensor	Trillium 120 QA
Sensitivity(V/m/s)	1200
Natural Period	120 Sec
Normalization Factor	831871100000000000
Normalization Frequency	1 Hz
No. Of Zeros	6
No. Of Poles	11
Poles & Zeros	
Zeros (rad/sec)	0
	0
	-31.63
	-160
	-350
	-3177
Poles (rad/sec)	$-0.036614 \pm 0.037059i$
	-32.55
	-142
	$-364 \pm 404i$
	-1260
	$-4900 \pm 5200i$
	$-7100 \pm 1700i$