

Instrument Response Information

Broadband Sensor	Reftek 151 120 A
Sensitivity (V/m/s)	2000
Natural Period	120 Sec
Normalisation Factor	60231900
Normalisation Frequency	1 Hz
No. of Zeros	2
No. of Poles	4
Poles and Zeros	
Zeros (rad/sec)	0
	0
Poles (rad/sec)	-0.03702 +0.03702 j
	-0.03702 -0.03702 j
	-222.14+222.14j
	-222.14-222.14j