

Instrument Response Information

Broadband Sensor	Trillium 240
Sensitivity (V/m/s)	1500
Natural Period	240 Sec
Normalisation Factor	362900
Normalisation Frequency	1 Hz
No. of Zeros	5
No. of Poles	7
Poles and Zeros	
Zeros (rad/sec)	0
	0
	-90
	-164.2
	-3203
Poles (rad/sec)	$-0.01813 + 0.01803j$
	$-0.01813 - 0.01803j$
	-99.9
	$-197.5 + 256.1j$
	$-197.5 - 256.1j$
	$-569 + 1150j$
	$-569 - 1150j$