

Instrument Response Information

Broadband Sensor	Trillium 120P
Sensitivity(V/m/s)	1201
Normalization Factor	1695000000
Normalization Frequency	1 Hz
No. Of Zeros	4
No. Of Poles	7
Poles & Zeros	
Zeros (rad/sec)	0
	0
	-106
	-158
Poles (rad/sec)	-0.03859 +0.03649i
	-0.03859 -0.03649i
	-190
	-158 +193i
	-158 -193i
	-639 +1418i
	-639 -1418i