

## Instrument Response Information

Broadband Sensor	Trillium Compact
Sensitivity(V/m/s)	750
Normalization Factor	818400000000
Normalization Frequency	1 Hz
No. Of Zeros	3
No. Of Poles	7
Poles & Zeros	
Zeros (rad/sec)	0
	0
	-434.1
Poles (rad/sec)	-0.03691+0.03712i
	-0.03691-0.03712i
	-371.2
	-373.9+475.5i
	-373.9-475.5i
	-588.4+1508i
	-588.4-1508i